FRESENIUS KABI

***** **CERTIFICATE OF ANALYSIS** *****

Article:	CompoStop Flex2F PLT processing and leuko red.syst		
Article number:	FD51600		
Batch number:	41TI28GA00	Quantity:	15 480
Manufacturing date:	23 September 2022	Expiry date:	2024-08

Test	**Limits**	**Results**
Configuration	Conform product specification	Conform
Physical control	In accordance with specifications	Conform
Bacterial Endotoxines	< 20 EU/device	Conform
Sterility	Sterile (Ph.Eur.)	Conform
Used Filterbatch	22PB007	

Batch released by Quality Assurance.		Manufactured by:
Date:	20.10.2022	Fresenius Kabi AG Plant: Fresenius Kabi Horatev CZ s.r.o. Hořátev 104, 289 13 Czech Republic
Name:	Klímová Vladimíra	
Signature:	V. Klímová	

Form.no. 17-4269/2